



DIVISION INSTRUMENTS
 81, Boulevard Jean-Baptiste Oudry, 94000 Créteil, France

Regulation sheet – Power supplies

The purpose of this regulatory fact sheet is to detail Essilor’s minimum requirements in terms of the industrial standards and regulations with which the Supplier undertakes to comply when supplying Products, in accordance with Essilor’s General Terms and Conditions of Purchase. In all cases Supplier undertakes to comply with all industrial standards and regulations applicable to the Products, even where not listed in this fact sheet.

This fact sheet may be revised by Essilor at any time, most notably following legislative or regulatory changes.

Supplier guarantees that the Products are compliant with the provisions of RoHS directive 2011/65/EU and all later related, modificatory or supplementary instruments in force at the time of delivery.

With a view to the Chinese market, Supplier shall submit a test report compliant with standard SJ / T11365 and a document indicating for each hazardous substance listed in standard SJ / T11363 its location and weight in the Product, in addition to the recycling percentage.

Pursuant to Article 33 para. 1 of REACH regulation 1907/2006, Supplier shall inform Essilor of the presence of candidate substances for authorization above a concentration of 0.1% weight by weight as a ratio to total weight in the products supplied, basing such notification on the “Candidate List” in force at the time of delivery. Supplier shall also provide this information within 30 days of the publication of updates to the “Candidate List”.

Supplier shall provide sufficient information available to Supplier to allow safe use of the Product concerned.

Supplier guarantees the compliance of the power supply with standard UL 61010-1 or UL 62368-1. If the power supply is used in Medical Devices in accordance with Regulation 2017/745/EU, Supplier guarantees that the power supply is compliant with standard UL 60601-1.

Supplier shall provide the electrical characteristics, the test report per the aforementioned standard along with all supporting evidence of the UL compliance of components considered hazardous by Essilor’s official approval body.

In the event of a decision to modify the composition of the Product, most notably following the adoption of regulatory provisions requiring authorization or restriction for a substance present in the article, Supplier shall inform Essilor 60 days prior to the implementation of the change. Supplier shall communicate to Essilor the elements listed above as modified following the change to Product composition.

In case of problems of interpretation between any of the headings appearing above the clauses and any of the corresponding clauses, the headings shall be considered as non-existent.

Departments to receive the above information:

Purchase Department : goutania@essilor.fr

Quality Department : qualite.instruments@essilor.fr